

JAN 05 2010

VERIFICATION OF CALIBRATION REPORT
of DataMaster edm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster edm S/N <u>130121</u> ✓
Name:	<u>John Pernell</u>	ID# <u>6260</u> Date <u>12/31/09</u> ✓
Agency:	<u>Fairbanks Airport Police</u> Phone # <u>474-2555</u> ✓	
Instrument Location:	<u>5195 Brumbaugh Blvd. Fairbanks, AK 99709</u> ✓	
A		
Also S/N	<u>X124670</u> ✓	Target Value <u>.082</u> High Pressure <u>1250</u>
B		
Also Test Values:	<u>.081</u> ✓ 1 st Also	<u>.080</u> ✓ 2 nd Also
Signature:	<u>John Pernell</u> ✓	
(OVER)		

BMB
1/13/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of January 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



